



# Certificate of Traceability

## Level I Laboratory Certification

This certifies that **PPD Global Central Laboratories**, using **Tosoh G8** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#9**

The system evaluated was:

Instrument: <b>G8</b>	Calibrator Lot: <b>ZS7001, ZS7001</b>	Column Lot: <b>N</b>
Reagent Lot: <b>N8-108D, N8-208D, N8-306D, HW-26D</b>	Calibrator Assigned Values: <b>5.70%, 10.90%</b>	

Date of Certification: May 1, 2019

Certification Expires: May 1, 2020

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor